

DABIT-013
Gallic Acid Reagent
Ecstasy (MDMA) / MDVP (Bath Salts) Test

www.fieldforensics.com



1) Remove Sampler from clear tube.

2) Remove paper tab from **Sticky Sampler Tip**. **DO NOT TOUCH** sticky sampler tip.



3) **Dab** sticky Sampler Tip on the suspect material. **Dab firmly** to collect a sample of suspect material.

4) **Return Sampler** to Clear Inner Tube, Sampler tip first, and close.



5) Hold Sampler as shown. **Squeeze Clear Inner Tube AND Sampler Tube** to break ampoule.



6) Keeping the **Sampler Tip** pointed downward, look for a color change on the **Sampler Tip** and compare to the color chart below. Observe for 60 seconds; if no color then the test is negative.

7) After recording results, carefully open clear tube and **SLOWLY** add five to six drops of the Neutralizer liquid (provided). Close cap on clear tube.

Possible Irritant:

Avoid contact of liquids with eyes and mucous membranes.

September 2015 Rev. 9

FIELD FORENSICS INC.

DABIT-013

Gallic Acid Reagent
MDMA / MDPV Test

Negative Test

MDPV (Bath Salts)

Ecstasy (MDMA)

Full DABIT™ Color Chart Available at www.fieldforensics.com